

NOTA DINAS
NOMOR : 017/RSMU/PM/XI/2022

Yth : Zainal Arifin, A.Md.PK (GM PENUNJANG MEDIS)
Dari : Otto Desyanto, S.Kep
Hal : Permohonan Penggantian All in One (Monitor PC)USG
Sifat : PENTING
Tanggal Nota : 08-11-2022

Isi :

Assalamualaikum wr. wb.

Sehubungan dengan trouble All in One (Monitor USG) yang tidak berfungsi disebabkan windows bermasalah dengan monitor Apple yang terkunci oleh password , maka kami mengajukan pembelian monitor all in one agar alat USG dapat berfungsi kembali (service report dan spesifikasi minimal all in one terlampir). Adapun perkiraan harga all in one adalah sekitar Rp. 11.500.000,- (sebelas juta lima ratus ribu rupiah). Demikian permohonan ini kami buat, atas perhatiannya kami ucapkan terima kasih.

Wassalamualaikum wr. wb.

Hormat Kami,
MANAGER MANAGER PENUNJANG
DIAGNOSTIK DAN TERAPI

Otto Desyanto, S.Kep

Jawaban :

Zainal Arifin, A.Md.PK (09-11-2022)

Waalaikumussalam wr. wb.

sehubungan dengan permasalahan tersebut, sudah seharusnya untuk disediakan komputer dengan spesifikasi terlampir. Mohon berkoordinasi dengan Instalasi TI



PT. SURYA TAMA MEDIKA
MEDICAL HOSPITAL SUPPLY • REPAIR AND SERVICE

SERVICE REPORT

NO: 2 NOVEMBER 2022.

Customer / Hospital : Rumah Sakit Pata Urdaan
Address : Jl. urdaan kulon NO:19 Surabaya.

Equipment Information

Product Name : ANISO Brand : Quantel Medical.
Serial Number : 1922 Status : Warranty

No Warranty

Problem Identification

Complaint / Problem : -PC ALL IN ONE TIDAK BISA
- Windows Bermasalah Tdk bisa Ext dg MAC.

Problem Cause / Solution

: -Disarankan ganti PC ALL IN ONE
- Untuk Spesifikasi terlampir.

Part Exchange

No.	Part Number	Description	Qty

Notes :

Engineer :

Mada. G.

Customer :

4.6. Computer minimum requirement

Processor	Intel® Pentium® IV Processor 2 GHz or higher Intel® Pentium® Centrino® Processor 1,6 GHz or higher Intel® Core2 Duo® Processor 1,8 GHz or higher AMD® Athlon®, Athlon Fx®, Athlon 64®, Athlon 64 X2® 2000+ or higher
Screen format	Minimum resolution desktop: 1680 x 1050 (WSXGA+) Minimum resolution laptop: 1680 x 1050 (WSXGA+) 16/9 or 16/10 format screen is recommended. (5/4 format screen is not recommended) Do not use the WXGA (1440x900) format. This format is not compatible some functions of the AVISO software will not be displayed: cross vector function, category...
Recommended monitors	15.4" WUXGA+ (1920 x 1200) or (1320 x 1080) 20" WSXGA+ (1680 x 1050)
Operating system	Microsoft® Windows® Seven (32 bits) Microsoft® Windows® Seven (64 bits) Microsoft® Windows® Eight (64 bits): 8.1 version (this version is required to display the AVISO application (desktop) by default upon startup)
Memory	Minimum 2Go DDR SDRAM (4Go DDR SDRAM recommended)
Graphics card	128MB RAM hardware compatible with DIRECTX® 9.0 (pixel shalder version 2.0) or higher <i>Example of compatible graphics card:</i> <i>NVIDIA® GeForce® 6600 or higher and NVIDIA® QuadroFX®</i> <i>ATI® Radeon® 9600 or higher ("X" letter corresponding to 10. Example ATI® Radeon® X300 corresponding to ATI® Radeon® 10300)</i> Integrated graphics chipset or memory shared graphic card are not compatible with our software for the moment (by example Intel® graphics)
Boot hard drives	60Go ATA-100 IDE (5,400 tpm) minimum
Computer and peripherals	3 USB 2.0 minimum. The computer has to be in conformity with the IEC 60950 standard / the peripherals connected to the computer also have to be in conformity with the IEC 60950 standard.



WARNINGS

- To upgrade the system and/or for computer replacement and/or for Windows settings (such as UAC and administrator rights): contact **QUANTEL MEDICAL** Service Department or your local distributor.
- **It is not possible to upgrade the AVISO software from a computer with an architecture of 32 bits to a computer with an architecture of 64 bits.**
- **The upgrade from AVISO to AVISO S is impossible (et vice versa)**